



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ7C00AZ

TECHNOTHROMBIN[®] TGA RB 5x0.5mL

REF 5006209

LOT 7C51B00.02

31-Jul-2017

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RB Reagenz	7C51B00	31-Jul-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00)	140 - 417 259.8 OK

Durchführung von
Prepared by

Fikret NASUFI

23. SEP. 2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA

23. SEP. 2016

Datum / Date: