
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3V00AZ.01
---	---	-------------

TECHNOVIEW Arixtra CONT H 6x1mL

REF 5090014
LOT 3V51B00.02
 31-Aug-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.62 – 1.04 µg/mL	0.83	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.62 – 1.04 µg/mL	0.89	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Christoph ENNE

04. NOV. 2016

Datum / Date:

**Freigabe von
Approved by**


Stefan STEINER

04. NOV. 2016

Datum / Date: