
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Y00AZ.01
---	---	-------------

TECHNOVIEW UFH CAL 5x1mL

REF 5090070
LOT 0Y71B00.01
 30-Sep-2018

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW UFH Cal 1	5F13B03	30-Sep-2018
1x1mL	TECHNOVIEW UFH Cal 2	5G13B03	30-Sep-2018
1x1mL	TECHNOVIEW UFH Cal 3	5H13B03	30-Sep-2018
1x1mL	TECHNOVIEW UFH Cal 4	5J13B03	30-Sep-2018
1x1mL	TECHNOVIEW UFH Cal 5	5K13B03	30-Sep-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	TECHNOVIEW UFH Cont low LOT 3C13B03	0.23 – 0.43 IU/mL	0.27	OK
	TECHNOVIEW UFH Cont high LOT 3J13B03	0.69 – 1.16 IU/mL	0.86	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lucas ROSENMAYR

02. FEB. 2017

Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

02. FEB. 2017

Datum / Date: