
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0600.01
---	---	-----------

TECHNOVIEW® Dabigatran CAL Set 4x1mL

REF 5090210
LOT 0662B00.01
 31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Dabigatran Cal 1	1T45B00	31-Oct-2018
1x1mL	TECHNOVIEW Dabigatran Cal 2	1U45B00	31-Oct-2018
1x1mL	TECHNOVIEW Dabigatran Cal 3	1V45B00	31-Oct-2018
1x1mL	TECHNOVIEW Dabigatran Cal 4	1W45B00	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Dabigatran Cont low LOT 3742B01	98.9 – 183.7 ng/mL	156.0	OK
	TECHNOVIEW Dabigatran Cont high LOT 3645B00	234.4 – 390.6 ng/mL	354.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas ROSENMAYR

11. JAN. 2017

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

11. JAN. 2017

Datum / Date: