

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3600.01
---	---	-----------

TECHNOVIEW Dabigatran CONT High 5x1mL

REF 5090212
LOT 3642B01.01
 31-Dez-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		236.7 – 394.5 ng/mL	315.6	OK
Verwendbarkeit Usability Ceveron® alpha	Technoclot DTI	236.7 – 394.5 ng/mL	336.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Monika WEINBAUER

15. FEB. 2016
Datum / Date:

Freigabe von
Approved by


Lucas ROSENMAYR

15. FEB. 2016
Datum / Date: