

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3700.01
---	---	-----------

**TECHNOVIEW Dabigatran CONT Low 5x1mL**

**REF**      5090214  
**LOT**      3742B01.02  
      31-Dez-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		98.9 – 183.7 ng/mL	141.3	OK
Verwendbarkeit Usability Ceveron® alpha	Technoclot DTI	98.9 – 183.7 ng/mL	119.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
Lucas ROSENMAYR

09. MRZ. 2016  
Datum / Date:

**Freigabe von  
Approved by**

  
Michael KAFKA

09. MRZ. 2016  
Datum / Date: