



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2K00.02

FVIII Inhibitorplasma HCV neg. 5x1mL

REF 5159010
LOT 2K61B00.01
⌚ 30-Nov-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	FVIII Inhibitor clotting, Daptin TC	27.6 – 46.0 BE/mL	43.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas ROSENMAYR

13. DEZ. 2016

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

13. DEZ. 2016

Datum / Date: