




**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRB00AZ

TECHNOZYM® vWF:CBA ELISA Kit 96T

REF 5450301
LOT RB6BB00.01
 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	vWF:CBA Plate Collagen Typ III	RNC65B0	31-Oct-2018
1 x 0.3mL	CBA Conjugate	RNE61B0	31-May-2018
1x 80mL	Wash buffer Concentrate	RNA63B0	31-Aug-2018
1 x 90mL	Incubation buffer	RNB63B0	31-Aug-2018
1 x 12mL	TMB Substrate	RKN61B0	31-May-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.08	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.63 – 1.05 IU/mL	0.91	OK
	Low Control (BND61B0)	0.13 – 0.29 IU/mL	0.20	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

04. NOV. 2016

Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

04. NOV. 2016

Datum / Date: