



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQR100AZ

**TECHNOZYM® vWF:CBA Collagen Typ VI Kit 96T**

**REF** 5450321

**LOT** R163B00.01

**⌚** 30-Apr-2018

Inhalt Content	Reagenz Reagent	<b>LOT</b>	<b>⌚</b>
1 x 12	vWF:CBA Plate Collagen Typ VI	RNN61B0	31-Aug-2018
1 x 0.3mL	CBA Conjugate	RNH51B1	30-Jun-2018
1x 80mL	Wash buffer Concentrate	RNA62B0	30-Jun-2018
1 x 90mL	Incubation buffer	RNB62B0	31-May-2018
1 x 12mL	TMB Substrate	RKN61B0	31-May-2018
1 x 12mL	Stop solution	RKK61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.30	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.51 – 0.85 IU/mL	0.63	OK
	Low Control (BND61B0)	0.11 – 0.25 IU/mL	0.15	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

01. SEP. 2016

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

01. SEP. 2016

Datum / Date: