

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6PT00AZ
---	---	-----------

TECHNOZYM® ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461
LOT 6PT51B0.02
 31-Dec-2017

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.18	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	70.21	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	10.39	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

09. FEB. 2017

Datum / Date:

**Freigabe von
Approved by**

Christoph Enne
Christoph ENNE

09. FEB. 2017

Datum / Date: