



## Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2018-10	LOT #: N1166469
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>Coatest SP4 Factor VIII Kit (82409463)</b>					
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0098	PASS
Intercept	--	0.250	--	0.139	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.052	PASS
Intercept	--	0.406	--	0.188	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl <sub>2</sub>	N1166215	2019-11	(1x) 6 ml
CT SP Substrate	N1066060	2018-10	(1x) 12 ml
Factor Reagent	N1065817	2019-04	(4x) 3 ml
CT SP Buffer, pH 7.3	N1166252	2018-11	(1x) 20 ml
CT SP Phospholipid	N1166468	2018-11	(1x) 2 ml

PREPARED BY: NPATEL N.m. Patel DATE: 12/15/16 12-15-16

APPROVAL SIGNATURE: DTENCZA D. Dencza DATE: 12/15/16 12/15/16