

ASSAYS

M30 Apoptosense® ELISA
Prod. No. 10011

M65® ELISA
Prod. No. 10020

M65 EpiDeath® ELISA
Prod. No. 10040

M30 CytoDeath™ ELISA
Prod. No. 10900

ANTIBODIES

M30 CytoDEATH™

- Unconjugated Prod. No. 10700
- Biotin Prod. No. 10750
- Fluorescein Prod. No. 10800
- Orange Prod. No. 10830

M5 Keratin 18
Prod. No. 10600

M6 Keratin 18
Prod. No. 10650

Certificate of Analysis

M65® ELISA

REF 10020

9057-11



Vivalavida AB, Håstholmsvägen 32, 131 30 Nacka, Sweden
Phone: +46 8 122 053 00 • www.vlvbio.com • info@vlvbio.com

M65® ELISA

REF 10020

Kit Lot No. LOT PE-TS 3169	Expiry 2017-11
--	------------------------------

Component	A _{450 nm}	LOT	
M65 Standard A	0 U/L	0.057	P-TS 3170 2017-11
M65 Standard B	125 U/L	0.142	P-TS 3171 2017-11
M65 Standard C	250 U/L	0.276	P-TS 3172 2017-11
M65 Standard D	500 U/L	0.553	P-TS 3173 2017-11
M65 Standard E	750 U/L	0.842	P-TS 3174 2017-11
M65 Standard F	1 200 U/L	1.329	P-TS 3175 2017-11
M65 Standard G	2 000 U/L	2.150	P-TS 3176 2017-11
M6 Coated Microstrips			P-SS 1879 2018-05
M65 HRP Conjugate			P-TS 3180 2018-02
M65 Conjugate Dilution Buffer			P-TS 3179 2018-06
TMB Substrate			P-SS 2272 2018-11
Stop Solution			P-TS 1034 2019-01
Wash Tablet			P-TS 1846 2019-04
Controls	Expected value (U/L)	Obtained value (U/L)	LOT
M65 Control Low	160 – 240	210	P-TS 3177 2017-11
M65 Control High	850 – 1 150	885	P-TS 3178 2017-11

Test date: 2016-11-24	Sign: Gordana Pjelic Beba
Approved by QA: 2016-11-29	Sign: Anne Kuryle