



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1P00AZ.01

Coagulation Control N 5x1mL

REF 5020040  
LOT 1P37B01.01  
⌚ 30-Jun-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	ATIII Chromogenes Substrat: F IIa	72.8 – 121.3 %	120.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Christoph ENNE

10. MRZ. 2017  
Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

10. MRZ. 2017  
Datum / Date: