




ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3P00AZ.03

Coagulation Control A, 5x1mL

REF 5021055

LOT 3P38B01.01

 30-Jun-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	Siron LS, sec	62.3 – 76.9	69.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
Christoph ENNE

19. DEZ. 2016

Datum / Date:

Freigabe von  
Approved by

  
Stefan STEINER

19. DEZ. 2016

Datum / Date: