

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3D00.01
---	---	-----------

**TECHNOVIEW® Apixaban CONT High 5x1mL**

**REF**            5090270  
**LOT**            3D61B00.02  
        30-Apr-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		1.69 – 2.81 IU/mL	2.25	OK
		201.2 – 335.4 ng/mL	268.3	
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	201.2 – 335.4 ng/mL	250.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

*Lucas ROSENMAYR*

19. DEZ. 2016

Datum / Date:

**Freigabe von**  
**Approved by**

*Christoph ENNE*

19. DEZ. 2016

Datum / Date: