

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3900.01
---	---	-----------

TECHNOVIEW® Apixaban CONT Low 5x1ml

REF 5090271
LOT 3961B00.02
 30-Apr-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.57 – 1.05 U/mL	0.81	OK
		74.1 – 137.5 ng/mL	105.8	
Verwendbarkeit Usability Ceveron © alpha	Anti-Xa Apixaban	74.1 – 137.5 ng/mL	122.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lucas ROSENMAYR

19. DEZ. 2016
Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

19. DEZ. 2016
Datum / Date: