

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0F00AZ.03
---	---	-------------


F VIII Inhibitor Reagent Kit 2-4T.

REF 5152005
LOT 0F71B00.01
 31-Jul-2018

Inhalt Content	Reagenz Reagent	LOT	
2x2.92mL	FVIII normal Plasma	1G62B00	30-Nov-2018
1x1mL	FVIII Inhibitor Plasma	2K52B01	31-Jan-2019
1x1mL	FVIII Inhibitor Free Plasma	2M62B0D	31-Jul-2018
1x17mL	Imidazol Puffer	9J63BA0	31-Jul-2021

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability FVIII clotting, Dapttin TC Ceveron® alpha	FVIII Inhibitor Plasma LOT 2K52B01	13.5 – 22.5 BE/mL	21.2	OK
	FVIII Inhibitor Free Plasma LOT 2M62B0D	< 0.4 BE/mL	< 0.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lucas ROSENMAYR

17. MRZ. 2017
Datum / Date:

**Freigabe von
Approved by**


Suzana GRESAK PODAKAR

17. MRZ. 2017
Datum / Date: