



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1R00.01

Coagulation Reference, 5x1mL

REF 5220110

LOT 1R51B01.04

🕒 28-Feb-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	FIX CCN LOT 1P37B01	80.3 – 133.8 %	125.8	OK
	FIX CCA LOT 3P38B01	15.8 – 26.3 %	20.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Christoph ENNE

27. JAN. 2017  
Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

27. JAN. 2017  
Datum / Date: