




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ7PT00AZ

TECHNOZYM[®] ADAMTS-13 INH Control Set (2x0.5mL)

REF 5450463

LOT 7PT52B0.02

 31-Dec-2017

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.31	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	58.93	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	8.26	OK

Durchführung von
Prepared by

Michael KAFKA 

23. FEB. 2017

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

23. FEB. 2017

Datum / Date: