



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRR00AZ

**TECHNOZYM® ADAMTS-13 Antigen ELISA Kit 96T.**

**REF** 5450601

**LOT** RR71B00.01

30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 Ag Plate	RRS62B0	30-Sep-2018
1 x 0.3mL	Adamts13 Ag Conjugate	RQV32B2	30-Apr-2018
1x 80mL	Wash buffer Concentrate	RNA64B0	31-Dec-2018
1 x 90mL	Incubation buffer	RNB71B0	31-Jan-2019
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT61B0	30-Nov-2018
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT61B0	31-Dec-2018
1 x 0.5mL	Adamts13 Ag High Control	RRY61B0	31-Dec-2018
1 x 0.5mL	Adamts13 Ag Low Control	RRZ61B0	31-Dec-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	0.97	OK
Verwendbarkeit Usability	High Control (RRY61B0)	0.47 – 0.79 IU/mL	0.69	OK
	Low Control (RRZ61B0)	0.13 – 0.29 IU/mL	0.21	OK

**Durchführung von  
Prepared by**

Fikret NASUFI *Fikret Nasufi*

24. FEB. 2017

Datum / Date:

**Freigabe von  
Approved by**

Michael KAFKA *[Signature]*

24. FEB. 2017

Datum / Date: