



Certificate of Analysis

PRODUCT:	CoaTest SP4 FVIII Kit	PRODUCT #:	82409463
EXP. DATE:	2018-12	LOT #:	N0177259
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0097	PASS
Intercept	--	0.250	--	0.128	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.041	PASS
Intercept	--	0.406	--	0.154	PASS
Standard Curve (r)	0.99	--	--	0.996	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N1166598	2019-11	(1x) 6 ml
CT SP Substrate	N0177162	2019-01	(1x) 12 ml
Factor Reagent	N1267002	2019-06	(4x) 3 ml
CT SP Buffer, pH 7.3	N1266893	2018-12	(1x) 20 ml
CT SP Phospholipid	N1266894	2018-12	(1x) 2 ml

PREPARED BY:

NPATEL N.M. Patel

DATE:

01/20/17 *1-20-17*

APPROVAL SIGNATURE:

D TENOZA P. Tenoga

DATE:

01/20/17 *1/20/17*

QCM0082409463-4, Rev. 00