

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2288, 25 mg Substrate	PRODUCT # : 82085239
EXP. DATE: 2019-03	LOT # : N0361544
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820852

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.7	Pass
FREE p-NITROANILINE	0	0.5	mol %	0.2	Pass

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2288	N0361544	2019-03	2 ml

PREPARED BY:

M. Oettinger

DATE:

4/4/2016

APPROVAL SIGNATURE:

D. Tencza

DATE:

4/4/2016