



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0Z00AZ.02

TECHNOCHROM[®] FXIII 3 x 3mL

REF 5360010
LOT 0Z71B00.01
⌚ 28-Feb-2018

Inhalt	Reagenz	LOT	⌚
3 x 3 mL	FXIII Activator Reagent	KTM1948100	28-Feb-2018
3 x 3 mL	FXIII Detection Reagent	KTM1948100	28-Feb-2018
3 x 3 mL	FXIII NADPH	KTM1948100	28-Feb-2018
3 x 1 mL	FXIII Inhibitor Reagent	KTM1948100	28-Feb-2018
1 x 6 mL	Stabilizer solution	KTM1948100	28-Feb-2018

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P37B01	60.8 – 101.3%
	Coagulation Control A Lot: 3P38B01	28.5 – 47.5%

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

27. MRZ. 2017

Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI

27. MRZ. 2017

Datum / Date: