



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8F00AZ

TECHNOTHROMBIN[®] TGA SUB 5x1.5mL

REF 5006235

LOT 8F63B00.01

⌚ 31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1.5mL	TGA Substrat	8F63B00	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z61B00	TGA RC High LOT 7F61B00	407 - 678	509.9	OK
		TGA RC Low LOT 7D62B00	176 - 294	293.4	OK
	TGA Control Low (nM Peak Thrombin) LOT 2863B00	TGA RC High LOT 7F61B00	63 - 147	141.4	OK
		TGA RC Low LOT 7D62B00	57 - 143	115.9	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

08. MRZ. 2017

Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR *Suzana Podakar*

08. MRZ. 2017

Datum / Date: