
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control High 5x1mL

REF 5090174
LOT 3351B01.01
 31-Jan-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HbsAg HIV 1.2 Ab HCV	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		242.6 – 404.3 ng/mL	323.4	
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa	242.6 – 404.3 ng/mL	342.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Christoph ENNE

07. MRZ. 2017
Datum / Date:

**Freigabe von
Approved by**


Stefan STEINER

07. MRZ. 2017
Datum / Date: