

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7D00AZ
---	---	----------

Ceveron[®] TGA RC Low 5x1 mL

REF 5006224
LOT 7D61B.03
 31-Mar-2018

Inhalt Content	Reagenz Reagent	LOT	
5x1 mL	Ceveron [®] TGA RC Low	7D61B	31-Mar-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B65B00	176 – 293	276.5	OK
	TGA Control High Lag time (min) LOT 1B65B00	2.3 – 3.9	2.8	OK
	TGA Control High (AUC) LOT 1B65B00	1973 – 3288	2901.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

16. FEB. 2017
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *M. Kafka*

17. FEB. 2017
Datum / Date: