

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ8E00AZ
---	---	----------

## Ceveron® TGA SUB 5x3 mL

**REF**      5006237  
**LOT**      8E61B00.02  
      28-Feb-2018

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
5x3 mL	Ceveron® TGA SUB	8E61B00	28-Feb-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls <b>LOT</b> 1B65B00	Peak Thrombin nM	TGA RCH f.Ceveron <b>LOT</b> 7F52B	363 - 605	502.3	OK
		TGA RCL f.Ceveron <b>LOT</b> 7D61B	176 - 293	276.8	OK
	Lag time min	TGA RCH f.Ceveron <b>LOT</b> 7F52B	1.4 - 2.4	1.9	OK
		TGA RCL f.Ceveron <b>LOT</b> 7D61B	2.3 - 3.9	3.0	OK
	AUC nM	TGA RCH f.Ceveron <b>LOT</b> 7F52B	2072 - 3453	2872.2	OK
		TGA RCL f.Ceveron <b>LOT</b> 7D61B	1973 - 3288	2798.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von  
Prepared by

Fikret NAŞUFI *Fikret Nasufi*

22. MRZ. 2017

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *MK*

22. MRZ. 2017

Datum / Date: