




| | | |
|---|---|-----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ0300.01 |
|---|---|-----------|

Technoview® Rivaroxaban CAL Set High 5x1mL

REF 5090171
LOT 0363B00.03
 31-Jan-2018

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|----------------------------------|------------|---|
| 1x1mL | Technoview Rivaroxaban CAL1 High | 5161B00 | 31-Jan-2018 |
| 1x1mL | Technoview Rivaroxaban CAL2 High | 5261B00 | 31-Jan-2018 |
| 1x1mL | Technoview Rivaroxaban CAL3 High | 5361B00 | 31-Jan-2018 |
| 1x1mL | Technoview Rivaroxaban CAL4 High | 5411B03 | 30-Jun-2018 |
| 1x1mL | Technoview Rivaroxaban CAL5 High | 5561B00 | 31-Jan-2018 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|--|----------------------------|--------------------|----|
| Verwendbarkeit Usability Ceveron® alpha | Rivaroxaban Control High LOT 3351B00 | 242.6 – 404.3 ng/mL | 335.7 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

Durchführung von
Prepared by

Lucas
Lucas ROSENMAYR

01. MRZ. 2017

Datum / Date:

Freigabe von
Approved by

Christoph
Christoph ENNE

01. MRZ. 2017

Datum / Date: