



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0N00.02

TECHNOCHROM® F VIII:C 40T.

REF 5344101

LOT 0N65B00.02

⌚ 30-Nov-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
2x2mL	Substrate for factor Xa 10µmol	8M62B00	30-Sep-2021
2x2mL	F VIII:C Reagenz A	7M61B00	31-Mar-2018
2x2mL	F VIII:C Reagenz B	7L41B01	31-Dec-2017
1x1mL	Reference Standard 1, F VIII:C	2W61B00	31-Mar-2018
1x1mL	Reference Standard 2, F VIII:C	2X61B00	31-Mar-2018
1x1mL	Reference Standard 3, F VIII:C	2Y41B02	31-Dec-2017
1x1mL	Reference Standard 4, F VIII:C	2Z51B0D	30-Nov-2017
2x8mL	FVIII:C Reaction buffer	9T62B00	31-Jul-2021
3x30mL	FVIII:C Dilution buffer	9Y51B00	31-Oct-2020

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B01	69.0 – 115.0%	82.7	OK
	CCA LOT 3P38B01	19.5 – 32.5%	22.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

Datum / Date: 10-Mar-2017

Freigabe von
Approved by

Christoph ENNE

Datum / Date: 10. MRZ. 2017