
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRA00AZ
---	---	----------

TECHNOZYM® vWF:Ag ELISA Kit 96T

REF 5450201

LOT RA71B00.01

 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	
1 x	vWF:Ag Plate	RNF61B0	31-Jul-2018
1 x 0.3mL	Ag Conjugate	RNH61B0	31-Oct-2018
1x 80mL	Wash buffer Concentrate	RNA64B0	31-Dec-2018
1 x 90mL	Incubation buffer	RNB71B0	31-Jan-2019
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.22	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.44 – 0.74 IU/mL	0.68	OK
	Low Control (BND61B0)	0.10 – 0.22 IU/mL	0.18	OK

Durchführung von
Prepared by

Michael KAFKA 

21. FEB. 2017

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

21. FEB. 2017

Datum / Date: