



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM<sup>®</sup> ADAMTS-13 INH ELISA Kit 96T

REF 5450401

LOT RP72B00.01

🕒 31-Dec-2017

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x 12	Adamts13 INH Plate	RPS72B0	31-Mar-2019
1 x 0.3mL	Adamts13 INH Conjugate	RPV61B0	31-Jul-2018
1x 80mL	Wash buffer Concentrate	RNA64B0	31-Dec-2018
1 x 90mL	Incubation buffer	RNB71B0	31-Jan-2019
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH High Control	RPY51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.41	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	68.02	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	11.85	OK

Durchführung von  
Prepared by

Michael KAFKA 

14. APR. 2017  
Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

18. APR. 2017  
Datum / Date: