
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQRP00AZ
---	---	----------

## TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF      5450451  
LOT      RP6DB0S.01  
      31-Dec-2017

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 6	Adamts13 INH Plate	RPS6BBS	30-Nov-2018
1 x 0.3mL	Adamts13 INH Conjugate	RPV61B0	31-Jul-2018
1x 80mL	Wash buffer Concentrate	RNA63B0	31-Aug-2018
1 x 90mL	Incubation buffer	RNB63B0	31-Aug-2018
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH High Control	RPY51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.35	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	67.41	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	11.26	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

19. JAN. 2017

Datum / Date:

**Freigabe von**  
**Approved by**

Lucas ROSENMAYR *Lucas*

19. JAN. 2017

Datum / Date: