



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ8UT0AZ.02

TECHNOZYM® ADAMTS-13 Activity Calibrator Set 6x0.5mL

REF 5450761
LOT 8UT71B0.01
⌚ 31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1x0.5mL	Adamts-13 Activity Calibrator 1	1UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 2	2UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 3	3UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 4	4UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 5	5UT63B0	31-Oct-2018
1x0.5mL	Adamts-13 Activity Calibrator 6	6UT63B0	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.04	OK
	Calibrator 6	< 0.2	0.03	OK
Verwendbarkeit / Usability	High Control (RUY63B0)	0.59 – 0.98 IU/mL	0.85	OK
	Low Control (RUZ63B0)	0.16 – 0.38 IU/mL	0.26	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

06. FEB. 2017
Datum / Date:

Freigabe von
Approved by

Lucas
Lucas ROSENMAYR

06. FEB. 2017
Datum / Date: