
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9600.01
---	---	-----------

Ceveron® Cleaning Solution 3% 25 mL

REF 9820300
LOT 9671B00.01
 31-Jan-2018

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	12.3	OK
Volumen Volume	25.0 – 26.5mL	25.0	OK
Sterilität Sterility	0 CFU/mL	0	OK

Verwendbarkeit, Fibrinogen, Clauss method	CCN LOT 1P37B01	2.52 – 3.78	3.62	OK
Usability, Fibrinogen, Clauss method Ceveron® alpha	CCA LOT 3P38B01	0.93 – 1.39	1.00	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by



Suzana GRESAK PODAKAR

03. FEB. 2017
Datum / Date:

Freigabe von
Approved by



Stefan STEINER

03. FEB. 2017
Datum / Date: