



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ...00AZ

TECHNOZYM[®] PAP Complex ELISA Kit 96T.

REF TC12060

LOT 120604B6.01

⌚ 30-Nov-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	PAP Plate	RGY63B0	30-Sep-2018
1 x 0.3mL	anti PAP POX	RBK61B0	30-Jun-2018
2 x 20mL	Wash buffer Concentrate	RBE52B0	30-Nov-2017
1 x 20mL	Dilution Buffer	RAF61B0	30-Jun-2018
1 x 12mL	TMB Substrate	RKN61B0	31-May-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	PAP-Calibrator 1	1BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 2	2BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 3	3BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 4	4BM61B0	30-Jun-2018
1 x 0.5mL	PAP-Calibrator 5	5BM61B0	30-Jun-2018
1 x 0.5mL	PAP high Control	ABM61B0	31-Jul-2018
1 x 0.5mL	PAP low Control	BBM61B0	31-Jul-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.38	OK
	Calibrator 5	OD: < 0.2	0.02	OK
Verwendbarkeit Usability	High Control (ABM61B0)	98.9 – 164.9 ng/mL	122.4	OK
	Low Control (BBM61B0)	13.0 – 30.2 ng/mL	21.0	OK

Durchführung von
Prepared by

Michael KAFKA 

21. OKT. 2016
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

21. OKT. 2016
Datum / Date: