
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6E00AZ.01
---	---	-------------

**Fibrinogen Reagent 120T. 5x5mL**

REF      5138080  
LOT      6E62BB0.02  
      31-Mar-2018

<b>Analytische Prozedur Analytical Procedure</b>	<b>Anforderung Requirement</b>	<b>Ergebnis Result</b>	
pH	7.0 – 8.0	7.9	OK
Restfeuchte Residual moisture	≤3%	1.31	OK


Verwendbarkeit Usability Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P37B00	2.52 – 3.78 g/L	3.23	OK
	CCA <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B01	0.93 – 1.39 g/L	1.03	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
**Lucas ROSENMAYR**

07. SEP. 2016  
Datum / Date:

**Freigabe von  
Approved by**

  
**Michael KAFKA**

07. SEP. 2016  
Datum / Date: