



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6C0BAZ.03

TECHNOCLOT® PT PLUS 5x4mL

REF 5003220

LOT 6C72BA0.01

⌚ 31-Mar-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Ceveron® alpha	CCN	PT Plus (%)	88.5 – 111.5%	92.4	OK
	LOT 1P37B01	PT Plus (INR)	0.85 – 1.07	1.00	OK
	CCA	PT Plus (%)	12.4 – 15.6%	13.9	OK
	LOT 3P38B01	PT Plus (INR)	2.30 – 2.90	2.68	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von  
Prepared by

Christoph ENNE

16. MAI 2017  
Datum / Date:

Freigabe von  
Approved by

Suzana GRESAK FODAKAR

16. MAI 2017  
Datum / Date: