



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ7F00AZ

TECHNOTHROMBIN<sup>®</sup> TGA RC High 5x0.5mL

REF 5006214

LOT 7F61B00.02

28-Feb-2018

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC High	7F61B00	28-Feb-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z61B00	407 - 678 479.8 OK
	TGA Control Low (nM Peak Thrombin) LOT 2863B00	63 - 147 104.7 OK

Durchführung von  
Prepared by

Fikret NASUFI

18. APR. 2017  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA

18. APR. 2017  
Datum / Date: