

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0U00AZ.02
---	---	-------------

AK-Calibrant 4x1mL

REF 5010004
LOT 0U51B00.02
 31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	AK Calibrant A	1N23B03	31-Oct-2018
1x1mL	AK Calibrant B	3L34B01	31-Oct-2018
1x1mL	AK Calibrant C	3M34B01	31-Oct-2018
1x1mL	AK Calibrant D	3N34B01	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	CCN LOT 1P37B00 Technoclot [®] PT Owren automated	79.7 – 100.4%	95.9	OK
	CCA LOT 3P38B00 Technoclot [®] PT Owren automated	23.0 – 29.0%	26.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas Rosenmayr
Lucas ROSENMAYR

10. APR. 2016

Datum / Date:

Freigabe von
Approved by

Monika Weinbauer
Monika WEINBAUER

10. APR. 2016

Datum / Date: