
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1P00AZ.01
---	---	-------------

Coagulation Control N 50x1mL

REF 5020050
LOT 1P37B01.01
 30-Jun-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	Siron LS	24.9 – 30.7 sec	26.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lucas ROSENMAYR

27. FEB. 2017

Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

27. FEB. 2017

Datum / Date: