



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRA00AZ

TECHNOZYM[®] vWF:Ag ELISA Kit 96T

REF 5450201

LOT RA72B00.01

⌚ 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:Ag Plate	RNF71B0	31-Mar-2019
1 x 0.3mL	Ag Conjugate	RNH61B0	31-Oct-2018
1x 80mL	Wash buffer Concentrate	RNA64B0	31-Dec-2018
1 x 90mL	Incubation buffer	RNB71B0	31-Jan-2019
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	vWF Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF Calibrator 5	5ND61B0	30-Apr-2018
1 x 0.5mL	vWF High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.14	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.44 – 0.74 IU/mL	0.57	OK
	Low Control (BND61B0)	0.10 – 0.22 IU/mL	0.16	OK

Durchführung von
Prepared by

Fikret NASUFI

12. MAI 2017

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

12. MAI 2017

Datum / Date: