



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRK00AZ

TECHNOZYM<sup>®</sup> vWF:CBA ELISA Collagen Typ I Kit 96T

REF 5450311

LOT RK72B00.01

⌚ 30-Apr-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:CBA Plate Collagen Typ I	RNM62B0	31-Dec-2018
1 x 0.3mL	vWF:Ag Conjugate	RNH61B0	31-Oct-2018
1x 80mL	Wash buffer Concentrate	RNA64B0	31-Dec-2018
1 x 90mL	Incubation buffer	RNB63B0	31-Aug-2018
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK62B0	30-Sep-2018
1 x 0.5mL	vWF:CBA Calibrator 1	1ND61B0	30-Apr-2018
1 x 0.5mL	vWF:CBA Calibrator 2	2ND61B0	30-Apr-2018
1 x 0.5mL	vWF:CBA Calibrator 3	3ND61B0	31-May-2018
1 x 0.5mL	vWF:CBA Calibrator 4	4ND61B0	30-Apr-2018
1 x 0.5mL	vWF:CBA Calibrator 5	5ND61B0	30-Apr-2018
1 x 0.5mL	vWF:CBA High Control	AND61B0	31-May-2018
1 x 0.5mL	vWF:CBA Low Control	BND61B0	31-May-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.18	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND61B0)	0.51 – 0.85 IU/mL	0.75	OK
	Low Control (BND61B0)	0.10 – 0.24 IU/mL	0.17	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

12. MAI 2017

Datum / Date:

Freigabe von  
Approved by

Lucas ROSENMAYR *Lucas*

15. MAI 2017

Datum / Date: