



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6PT00AZ

TECHNOZYM® ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461

LOT 6PT51B0.03

⌚ 31-Dec-2017

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B0	31-Dec-2017
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B0	31-Dec-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.25	OK
	Calibrator 5	< 0.1	0.05	OK
Verwendbarkeit Usability	High Control (RPY51B0)	54.32 – 81.48 U/mL	65.75	OK
	Low Control (RPZ51B0)	5.27 – 12.29 U/mL	11.84	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

26. APR. 2017

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

26. APR. 2017

Datum / Date: