




**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**


DQRU00AZ

**TECHNOZYM® ADAMTS-13 Activity ELISA Kit 96T.**

**REF** 5450701

**LOT** RU74B00.01

 31-Aug-2018

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	705	31-Aug-2018
1 x 12mL	Adamts13 Activity HRP Conj antibody	705	31-Aug-2018
1 x 53mL	Wash buffer concentrate	705	31-Aug-2018
1 x 30mL	Reaction buffer	705	31-Aug-2018
1 x 12mL	Color Reagent	705	31-Aug-2018
1 x 12mL	Stop solution	705	31-Aug-2018
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	705	31-Aug-2018
1 x	Sample Dilution MTP	705	31-Aug-2018
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity High Control	RUY63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Low Control	RUZ63B0	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.43	OK
	OD Cal 6	< 0.2	0.03	OK
	High Control (RUY63B0)	0.59 – 0.98 IU/mL	0.63	OK
	Low Control (RUZ63B0)	0.16 – 0.38 IU/mL	0.22	OK



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRU00AZ

Durchführung von  
Prepared by

Michael KAFKA

A handwritten signature in blue ink, appearing to read "Michael Kafka", written over a horizontal line.

27. APR. 2017

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI

A handwritten signature in blue ink, appearing to read "Fikret Nasufi", written over a horizontal line.

27. APR. 2017

Datum / Date: