

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRU00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 Activity ELISA Kit 96T.

REF 5450701
LOT RU74B00.02
 31-Aug-2018

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 12	Adamts13 Activity Strips	705	31-Aug-2018
1 x 12mL	Adamts13 Activity HRP Conj antibody	705	31-Aug-2018
1 x 53mL	Wash buffer concentrate	705	31-Aug-2018
1 x 30mL	Reaction buffer	705	31-Aug-2018
1 x 12mL	Color Reagent	705	31-Aug-2018
1 x 12mL	Stop solution	705	31-Aug-2018
2 x 6mL	ADAMTS-13 Activity vWF73 Substrate	705	31-Aug-2018
1 x	Sample Dilution MTP	705	31-Aug-2018
1 x 0.5mL	Adamts13 Activity Calibrator 1	1UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 2	2UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 3	3UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 4	4UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 5	5UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Calibrator 6	6UT63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity High Control	RUY63B0	31-Oct-2018
1 x 0.5mL	Adamts13 Activity Low Control	RUZ63B0	31-Oct-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	> 1	1.37	OK
	OD Cal 6	< 0.2	0.03	OK
	High Control (RUY63B0)	0.59 – 0.98 IU/mL	0.86	OK
	Low Control (RUZ63B0)	0.16 – 0.38 IU/mL	0.26	OK

	<p>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p>DQRU00AZ</p>
---	---	-----------------

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

19. MAI 2017
Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

19. MAI 2017
Datum / Date: