
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3U00AZ.01
---	---	-------------

TECHNOVIEW Arixtra Control L 6x1mL

REF 5090012
LOT 3U51B01.01
 31-Mai-2019


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.32 – 0.60 µg/mL	0.46	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.32 – 0.60 µg/mL	0.39	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lucas ROSENMAYR

19. JUNI 2017
Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

19. JUNI 2017
Datum / Date: