

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00AZ.02
---	---	-------------

TECHNOCHROM® FXIII 3 x 3mL

REF 5360010
LOT 0Z72B00.01
 30-Apr-2018

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	KTM1949100	30-Apr-2018
3 x 3 mL	FXIII Detection Reagent	KTM1949100	30-Apr-2018
3 x 3 mL	FXIII NADPH	KTM1949100	30-Apr-2018
3 x 1 mL	FXIII Inhibitor Reagent	KTM1949100	30-Apr-2018
1 x 6 mL	Stabilizer solution	KTM1949100	30-Apr-2018

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P37B01	60.8 – 101.3%
	Coagulation Control A Lot: 3P38B01	28.5 – 47.5%

Durchführung von
Prepared by

Michael KAFKA 

14. JUNI 2017

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

14. JUNI 2017

Datum / Date: