

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0Z00AZ.02
---	---	-------------

## TECHNOCHROM® FXIII 3 x 3mL

REF      5360010  
LOT      0Z72B00.02  
      30-Apr-2018

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	KTM1949100	30-Apr-2018
3 x 3 mL	FXIII Detection Reagent	KTM1949100	30-Apr-2018
3 x 3 mL	FXIII NADPH	KTM1949100	30-Apr-2018
3 x 1 mL	FXIII Inhibitor Reagent	KTM1949100	30-Apr-2018
1 x 6 mL	Stabilizer solution	KTM1949100	30-Apr-2018

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P37B01	60.8 – 101.3% 84.8
	Coagulation Control A Lot: 3P38B01	28.5 – 47.5% 38.9

Durchführung von  
Prepared by

Michael KAFKA 

14. JUNI 2017

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

14. JUNI 2017

Datum / Date: