



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401

LOT RP73B00.01

⌚ 31-Oct-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 INH Plate	RPS73B0	30-Apr-2019
1 x 0.3mL	Adamts13 INH Conjugate	RPV51B1	30-Nov-2018
1x 80mL	Wash buffer Concentrate	RNA71B0	30-Apr-2019
1 x 90mL	Incubation buffer	RNB71B0	31-Jan-2019
1 x 12mL	TMB Substrate	RKN62B0	31-Oct-2018
1 x 12mL	Stop solution	RKK71B0	31-May-2019
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH High Control	RPY51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Low Control	RPZ51B1	30-Apr-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.47	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY51B1)	54.32 – 81.48 U/mL	67.09	OK
	Low Control (RPZ51B1)	5.27 – 12.29 U/mL	11.25	OK

Durchführung von  
Prepared by

Michael KAFKA

19. JUNI 2017

Datum / Date:

Freigabe von  
Approved by

Lucas ROSENMAYR

19. JUNI 2017

Datum / Date: