




ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ6PT00AZ

TECHNOZYM[®] ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461

LOT 6PT71B0.01

 30-Apr-2019

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT51B1	30-Apr-2019
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT51B1	30-Apr-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.42	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY51B1)	54.32 – 81.48 U/mL	67.13	OK
	Low Control (RPZ51B1)	5.27 – 12.29 U/mL	10.14	OK

Durchführung von
Prepared by

Michael KAFKA 

19. JUNI 2017

Datum / Date:

Freigabe von
Approved by

Lucas ROSENMAYR 

19. JUNI 2017

Datum / Date: